



# ACCREDITATION CERTIFICATE

Issued under the authority of Bangladesh Accreditation Act, 2006  
by Bangladesh Accreditation Board (BAB), Ministry of Industries to

**CS LAB LIMITED**

**28/B, 29/A, 29/B, LEVEL-7, KAKRAIL**

**VIP ROAD, DHAKA-1000, BANGLADESH**

This is to certify that this  
**Calibration Laboratory**

is accredited in accordance with the international standard  
**ISO/IEC 17025:2017**

in respect of the associated scope, subject to the terms and  
conditions governing the relevant conformity assessment  
body (CAB) accreditation.

Certificate Number : **02.012.21**  
Accreditation Date : **15 February 2021**  
Date of Issuance : **15 February 2021**  
Date of Expiration : **14 February 2024**



  
**Md. Monwarul Islam**  
Director General

This certificate must be returned on request; reproduction must follow BAB guidelines. For the specific scopes to which this accreditation applies, please refer to the Directory of CABs at BAB website.

## SCOPE OF ACCREDITATION

(For Calibration Laboratory)

**CAB Name & Address:** CS LAB LIMITED  
28/B, 29/A, 29/B, LEVEL-7, KAKRAIL, VIP ROAD,  
DHAKA-1000, BANGLADESH

**Accreditation Standard:** ISO/IEC 17025:2017  
**Accreditation Date:** 15 Feb 2021

**Certificate Number:** 02.012.21  
**Issued on:** 15 Feb 2021

**Last Amended on:** NA  
**Valid until:** 14 Feb 2024

**Amendment no:** NA

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
<b>Field: Thermal (In Lab &amp; On site)</b>				
01	Temperature (Digital Thermometer)	CS-LAB-CP-03	(-) 40°C to 0°C	±0.09°C
			>0°C to 150°C	±0.1°C
			>150°C to 250°C	±0.2°C
			>250°C to 400°C	±0.4°C
<b>Field: Mechanical</b>				
02	Pressure (On site & In lab)	CS-LAB-CP-01	0 to 10 bar	±0.08 bar
			10 to 20 bar	±0.09 bar
03	Volume (Micro- pipette)	CS-LAB-CP-05	10 µl to 100 µl	±0.15 µl
			100 µl to 1000 µl	±0.45 µl
			1000 µl to 10000 µl	±3.00 µl

\*\*\*END\*\*\*



Quality Manager