



ACCREDITATION CERTIFICATE

Issued under the authority of Bangladesh Accreditation Act, 2006
by Bangladesh Accreditation Board (BAB), Ministry of Industries to

M/s. Reverie calibration lab

**Reverie Power& Automation Engineering Ltd, 5th and 7th Floor
Evergreen Plaza, 260/B, Tejgaon I/A, Dhaka-1208, Bangladesh**

This is to certify that this

Calibration Laboratory

is accredited in accordance with the international standard

ISO/IEC 17025:2017

in respect of the associated scope, subject to the terms and
conditions governing the relevant conformity assessment
body (CAB) accreditation.

Certificate Number : **02.014.22**
Accreditation Date : **06 March 2022**
Date of Issuance : **06 March 2022**
Date of Expiration : **05 March 2025**




Md. Monwarul Islam
Director General

This certificate must be returned on request; reproduction must follow BAB guidelines. For the specific scopes to which this accreditation applies, please refer to the Directory of CABs at BAB website.


SCOPE OF ACCREDITATION

(For Calibration Laboratory)

CAB Name & Address: M/s Reverie calibration lab
 Reverie Power & Automation Engineering Ltd, 5th and 7th Floor,
 Evergreen Plaza, 260/B, Tejgaon I/A, Dhaka-1208, Bangladesh.

Accreditation Standard:	ISO/IEC 17025:2017	Accreditation Date:	6 March 2022
Certificate Number:	02.014.22	Issued on:	6 March 2022
Last Amended on:	NA	Valid until:	5 March 2025
Amendment no:	NA		

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2) (to be expressed in ±)
Field: Electro-Technical				
Source (Fluke 5080A)				
1.	DC Voltage	RCL/DCV/S01 (In-house method)	10mV – 300mV	± 1%
			300mV- 3V	± 0.06%
			3V- 30V	± 0.06%
			30V-300V	± 0.1%
			300V-1000V	± 1%
2.	AC Voltage (45Hz)	RCL/ACV/S01 (In-house method)	10mV – 300mV	± 1%
			300mV- 3V	± 0.6%
			3V- 30V	± 0.1%
			30V-300V	± 0.1%
			300V-1000V	± 1%
	AC Voltage (1000Hz)	RCL/ACV/S01 (In-house method)	30mV – 300mV	± 1%
			300mV- 3V	± 0.19%
			3V- 30V	± 0.1%
3.	DC Current	RCL/DCI/S01 (In-house method)	3mA- 30mA	± 0.19%
			30mA- 300mA	± 1%




Quality Manager

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: M/s Reverie calibration lab
Reverie Power & Automation Engineering Ltd, 5th and 7th Floor,
Evergreen Plaza, 260/B, Tejgaon I/A, Dhaka-1208, Bangladesh.

Accreditation Standard:	ISO/IEC 17025:2017	Accreditation Date:	6 March 2022
Certificate Number:	02.014.22	Issued on:	6 March 2022
Last Amended on:	NA	Valid until:	5 March 2025
Amendment no:	NA		

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2) (to be expressed in ±)
			300mA- 3A	± 0.07%
			3A- 10A	± 0.08%
			1mA- 300mA	± 1%
			300mA- 3A	± 0.13%
4.	AC Current (45Hz)	RCL/ACI/S01 (In-house method)	3A-10A	± 0.23%
			10A-30A	± 1%
			30A-300A	± 1%
			300A-1000A	± 1%
	AC Current (1000Hz)	RCL/ACI/S01 (In-house method)	3mA- 300mA	± 0.2%
			300mA- 3A	± 0.1%
			3A-10A	± 0.1%
			1kΩ-100kΩ	± 0.06%
5.	Resistance	RCL/RES/S01 (In-house method)	100kΩ-1MΩ	± 0.1%
			1MΩ-100MΩ	± 0.1%
		Measure (Fluke 8588A)		
			300mV- 1V	± 0.001%



Quality Manager

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: M/s Reverie calibration lab
Reverie Power & Automation Engineering Ltd, 5th and 7th Floor,
Evergreen Plaza, 260/B, Tejgaon I/A, Dhaka-1208, Bangladesh.

Accreditation Standard:	ISO/IEC 17025:2017	Accreditation Date:	6 March 2022
Certificate Number:	02.014.22	Issued on:	6 March 2022
Last Amended on:	NA	Valid until:	5 March 2025
Amendment no:	NA		

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2) (to be expressed in ±)
6.	DC Voltage	RCL/DCV/M01 (In-house method)	1V- 10V	± 0.004%
			10V-100V	± 0.005%
			100V-1000V	± 0.005%
7.	AC Voltage (45Hz)	RCL/ACV/M01 (In-house method)	3mV- 100mV	± 0.40%
			100mV- 1V	± 0.33%
			1V-10V	± 0.04%
			10V- 100V	± 0.04%
			100V- 1000V	± 0.04%
	AC Voltage (8kHz)	RCL/ACV/M01 (In-house method)	500V-1000V	± 0.04%
	AC Voltage (10kHz)	RCL/ACV/M01 (In-house method)	100mV- 1V	± 0.32%
			1V-10V	± 0.04%
			10V- 100V	± 0.04%
8.	DC Current	RCL/DCI/M01 (In-house method)	1mA- 100mA	± 0.02%
			100mA-1A	± 0.03%
			1A-10A	± 0.05%



 Quality Manager

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: M/s Reverie calibration lab
Reverie Power & Automation Engineering Ltd, 5th and 7th Floor,
Evergreen Plaza, 260/B, Tejgaon I/A, Dhaka-1208, Bangladesh.

Accreditation Standard:	ISO/IEC 17025:2017	Accreditation Date:	6 March 2022
Certificate Number:	02.014.22	Issued on:	6 March 2022
Last Amended on:	NA	Valid until:	5 March 2025
Amendment no:	NA		

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2) (to be expressed in ±)
9.	AC Current (45Hz)	RCL/ACI/M01 (In-house method)	3mA- 100mA	± 0.06%
			100mA- 1A	± 0.06%
			1A- 10A	± 0.1%
			10A-20A	± 0.1%
	AC Current (5kHz)	RCL/ACI/M01 (In-house method)	3mA- 100mA	± 0.07%
			100mA- 1A	± 0.06%
			1A- 10A	± 0.1%
				± 0.07%
10.	Resistance	RCL/RES/M01 (In-house method)	50Ω - 1kΩ	± 1%
			1kΩ - 1MΩ	± 0.02%
			1MΩ - 200MΩ	± 0.05%
11.	Frequency	RCL/FR/M01 (In-house method)	45Hz- 100kHz	± 0.6%
Field: Thermal (Fluke 9132)				
1.	Temperature	RCL/TH/S01 (In-house method)	50°C - 500°C	±2.71 °C to ±4.2 °C



 Quality Manager