



## Bangladesh Accreditation Board (BAB)

91, Motijheel C/A, Dhaka-1000, Tel: +880-2-9513221 Fax: +880-2-9513222

Email: [info@bab.org.bd](mailto:info@bab.org.bd) Web: [www.bab.org.bd](http://www.bab.org.bd)

# Certificate of Accreditation

This is to certify that

**Resource Instrument & Measurement Enterprise (RIME)**

**Prabal Housing, House No: 05, Ring Road**

**Adabar, Mohammadpur, Dhaka**

has been granted accreditation in respect of the scope of accreditation described in the attached document, subject to the terms and conditions governing the relevant Conformity Assessment Body (CAB) accreditation

This Calibration Laboratory having met the requirements of ISO/IEC 17025:2005 and the BAB, is accredited for Thermal (Temp.) & Mechanical (Pressure) as described in the associated Scope of Accreditation.



**Certificate Number :** 02.006.17

**Issued on :** 27 December 2017

**Accreditation Date :** 27 December 2017

**Valid until :** 26 December 2020

**Director General**

**Bangladesh Accreditation Board (BAB)**

This certificate has been issued under the authority of Bangladesh Accreditation Act, 2006 and must be returned on request; reproduction must follow guidelines in place at date of issue. For the specific scopes to which this accreditation applies, please refer to the directory of accredited CABs at <http://www.bab.org.bd/directory-of-accredited-cabs>




## SCOPE OF ACCREDITATION (For Calibration Laboratory)

**CAB Name & Address:** Resource Instrument & Measurement Enterprise (RIME)  
 Prabal Housing, House No: 05, Ring Road  
 Adabar, Mohammadpur, Dhaka.

<b>Accreditation Standard:</b>	ISO/IEC 17025:2005	<b>Accreditation Date:</b>	27 December 2017
<b>Certificate Number:</b>	02.006.17	<b>Issued on:</b>	27 December 2017
<b>Last Amended on:</b>	-	<b>Valid until:</b>	26 December 2020
<b>Amendment no:</b>	-		

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
<b>Field: Thermal</b>				
01	Temperature	TK 112S Using PRT with Digital Read out & Dry Block Bath By Comparison Method (CP 02)	30 °C to 300 °C	±0.20 °C
<b>Field: Mechanical</b>				
02	Pressure	DG-10-S Using Digital Pressure Gauge with Pneumatic and Hydraulic Pressure Comparator By Comparison Method (CP 01)	0 Bar to 40 Bar	±0.2 Bar

\*End\*

  
 \_\_\_\_\_  
 Quality Manager