



Bangladesh Accreditation Board (BAB)

91, Motijheel C/A, Dhaka-1000, Tel: +880-2-9513221 Fax: +880-2-9513222

Email: info@bab.org.bd Web: www.bab.org.bd

Certificate of Accreditation

This is to certify that

Standard Calibration Service Private Limited

Rabeya Commercial Complex, 33/Kha (5th Floor), Section-06

Mirpur Circle-10, Dhaka-1216, Bangladesh

has been granted accreditation in respect of the scope of accreditation described in the attached document, subject to the terms and conditions governing the relevant Conformity Assessment Body (CAB) accreditation

This Calibration Laboratory having met the requirements of ISO/IEC 17025:2005 and the BAB, is accredited for Mechanical as described in the associated Scope of Accreditation.



Certificate Number : 02.009.18

Issued on : 13 September 2018

Accreditation Date : 13 September 2018

Valid until : 12 September 2021

Md. Monwarul Islam
Director General

Bangladesh Accreditation Board (BAB)

This certificate has been issued under the authority of Bangladesh Accreditation Act, 2006 and must be returned on request; reproduction must follow guidelines in place at date of issue. For the specific scopes to which this accreditation applies, please refer to the directory of accredited CABs at <http://www.bab.org.bd/directory-of-accredited-cabs>

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: Standard Calibration service Private Limited.

Rabeya Commercial Complex, 33/Kha (5th Floor) Section-6,
Mirpur Circle-10, Dhaka-1216, Bangladesh

Accreditation Standard: ISO/IEC 17025:2005

Accreditation Date: 13 Sep 2018

Certificate Number: 02.009.18

Issued on: 13 Sep 2018


Last Amended on: NA

Valid until: 12 Sep 2021

Amendment no: NA

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
Field: Mechanical				
01.	Mass Balance (Onsite & Lab)	SCS-CP-01	Max Cap 200 g Readability \geq 0.1 mg	3.00 mg
			Max Cap 500 g Readability \geq 1 mg	12 mg
			Max Cap 1 kg Readability \geq 10 mg	18 mg
			Max Cap 2 kg Readability \geq 10 mg	31 mg
			Max Cap 5 kg Readability \geq 0.1 g	0.114 g
			Max Cap 10 kg Readability \geq 0.1 g	0.17 g
			Max Cap 20 kg Readability \geq 0.1 g	0.34 g
			Max Cap 25 kg Readability \geq 0.1 g	0.41 g

END



Quality Manager