



ACCREDITATION CERTIFICATE

Issued under the authority of Bangladesh Accreditation Act, 2006
by Bangladesh Accreditation Board (BAB), Ministry of Industries to

EQMS Consulting Limited (Testing Laboratory)

Flat#F1, House#Ta-134/A, Boishakhi Sarani

Gulshan-Badda Link Road, Dhaka, Bangladesh. Post Code- 1212

This is to certify that this

Testing Laboratory

is accredited in accordance with the international standard

ISO/IEC 17025:2017

in respect of the associated scope, subject to the terms and
conditions governing the relevant conformity assessment
body (CAB) accreditation.

Certificate Number : 01.062.21
Accreditation Date : 26 August 2021
Date of Issuance : 26 August 2021
Date of Expiration : 25 August 2024




Md. Monwarul Islam
Director General

This certificate must be returned on request; reproduction must follow BAB guidelines. For the specific scopes to which this accreditation applies, please refer to the Directory of CABs at BAB website.

ACCREDITATION SCHEME

(For Testing Laboratory)

CAB Name & Address: EQMS Consulting Limited (Testing Laboratory), Flat#F1, House#Ta-134/A, Boishakhi Sarani, Gulshan-Badda Link Road, Dhaka, Bangladesh. Post Code- 1212

Accreditation Standard: ISO/IEC 17025:2017 **Accreditation Date:** 26 Aug 2021

Certificate Number: 01.061.21 **Issued on:** 26 Aug 2021

Last Amended on: **Valid until:** 25 Aug 2024

Amendment no:

S.N.	Products/ Materials/ Items of test	Type of tests performed	Specifications/ Standard test methods/Techniques used	Range of testing/Limit of detection
Field:				
1.	Drinking Water/Waste water/Surface Water	pH	SOP-03/ Hanna Combo Meter HI 98130; Ion-electrode method	0.00 to 14.00
2.	Drinking Water/Waste water/Surface Water	Temperature	SOP-03/ Hanna pH Meter HI 98130; Ion-electrode method	0.0 to 60.0°C or 32.0 to 140.0°F
3.	Drinking Water/Waste water/Surface Water	TDS	SOP-02/ Hanna EC Meter HI 99301 (EC, TDS & Temperature Meter); ion-electrode method	0.00 to 10.00 ppt
4.	Drinking Water/Waste water/Surface Water	EC	SOP-02/ Hanna EC Meter HI 99301 (EC, TDS & Temperature Meter); ion-electrode method	0.00 to 20.00 mS/cm
5.	Drinking Water/Waste water/Surface Water	Dissolved Oxygen	SOP-01/ DO-5509; ion-electrode method	0 to 20.0 mg/L
6.	Drinking Water/Waste water/Surface Water	Salinity	SOP-04/ AZ 8371; ion-electrode method	0.00-10.0 ppt (NaCl) & 10.0-70.0ppt

Field:

END



Quality Manager