



ACCREDITATION CERTIFICATE

Issued under the authority of Bangladesh Accreditation Act, 2006
by Bangladesh Accreditation Board (BAB), Ministry of Industries to

Peerless Calibration Services Ltd.

2F, 8C & 9C, Zoo Road, Section-1, Mirpur

Dhaka-1216, Bangladesh.

This is to certify that this
Calibration Laboratory

is accredited in accordance with the international standard
ISO/IEC 17025:2017

in respect of the associated scope, subject to the terms and
conditions governing the relevant conformity assessment
body (CAB) accreditation.

Certificate Number : **02.011.20**
Accreditation Date : **29 October 2020**
Date of Issuance : **23 May 2023**
Date of Expiration : **28 October 2026**




Md. Monwarul Islam
Director General


This certificate must be returned on request; reproduction must follow BAB guidelines. For the specific scopes to which this accreditation applies, please refer to the Directory of CABs at BAB website.

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: Peerless Calibration Services Ltd.
 2F, 8C & 9C, Zoo Road, Section-1, Mirpur, Dhaka-1216, Bangladesh.

Accreditation Standard: ISO/IEC 17025:2017 **Accreditation Date:** 29 October 2020
Certificate Number: 02.011.20 **Issued on:** 23 May 2023
Last Amended on: NA **Valid until:** 28 October 2026
Amendment no: NA

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
Field: Electrical				
1.	AC Current (10Hz to 1kHz) (Measure)	WI-ET-05	10 μ A to 100 μ A 100 μ A to 400 μ A 400 μ A to 4000 μ A 4 mA to 400 mA 100 μ A to 1 A 1 A to 3 A 3 A to 10 A	\pm 0.80 % to 0.06 % \pm 0.06 % to 0.04 % \pm 0.04 % to 0.25 % \pm 0.76 % to 0.18 % \pm 0.23 % to 0.17 % \pm 0.22 % to 0.19 % \pm 0.23 % to 0.25 %
	AC Current (45 Hz to 1 kHz) (Source)	WI-ET-05	0.2 mA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 1 A to 10 A	\pm 0.24 % to 0.34 % \pm 0.34 % to 0.36 % \pm 0.34 % to 0.36 % \pm 0.34 % to 0.17 % \pm 0.17 % to 0.27 %
2.	DC Current (Measure)	WI-ET-05	10 μ A to 100 μ A 100 μ A to 400 μ A 400 μ A to 4000 μ A 4 mA to 400 mA 100 μ A to 1 A 1 A to 3 A 3 A to 10 A	\pm 0.090 % to 0.020 % \pm 0.020 % to 0.005 % \pm 0.005 % to 0.006 % \pm 0.006 % to 0.011 % \pm 0.011 % to 0.053 % \pm 0.17 % to 0.09 % \pm 0.14 % to 0.09 %
	DC Current (Source)	WI-ET-05	0.2 mA to 2 mA 2 mA to 200 mA 200 mA to 2 A 1 A to 10 A	\pm 0.07 % to 0.18 % \pm 0.17 % to 0.09% \pm 0.09 % to 0.20 % \pm 0.10 % to 0.20 %
3.	AC High Current (10 Hz to 500 Hz) DC High Current (Measure)	WI-ET-02	10 A to 1000 A 10 A to 1000 A	\pm 0.180 to 0.099 % \pm 0.440 to 0.100 %
	AC High Current (50 Hz) DC High Current (Source)	WI-ET-02	10 A to 1000 A 10 A to 1000 A	\pm 1.3 % \pm 0.81 %




Quality Manager

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: Peerless Calibration Services Ltd.
 2F, 8C & 9C, Zoo Road, Section-1, Mirpur, Dhaka-1216, Bangladesh.

Accreditation Standard: ISO/IEC 17025:2017 **Accreditation Date:** 29 October 2020
Certificate Number: 02.011.20 **Issued on:** 23 May 2023
Last Amended on: NA **Valid until:** 28 October 2026
Amendment no: NA

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
4.	AC Voltage (10Hz to 1kHz) (Measure)	WI-ET-05	1 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 400 V 400 V to 750 V 400 V to 1000 V	± 0.471 % to 0.019 % ± 0.019 % to 0.008 % ± 0.008 % to 0.007 % ± 0.007 % to 0.008 % ± 0.009 % to 0.010 % ± 0.17 % to 0.10 % ± 0.17 % to 0.06 %
	AC Voltage (45 Hz to 1 kHz) (Source)	WI-ET-05	10 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 100 V to 1000 V	± 0.12 % to 0.20 % ± 0.17 % to 0.24 % ± 0.24 % to 0.26 % ± 0.24 % to 0.26 % ± 0.12 % to 0.11 %
5.	DC Voltage (Measure)	WI-ET-05	1 mV to 200 mV 200 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 400 V 400 V to 750 V 400 V to 1000 V	± 0.0480 % to 0.0014 % ± 0.0014 % to 0.0007 % ± 0.0009 % to 0.0007 % ± 0.0008 % to 0.0007 % ± 0.0007 % to 0.0009 % ± 0.14 % to 0.03 % ± 0.17 % to 0.04 %
	DC Voltage (Source)	WI-ET-05	10 mV to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 100 V to 1000 V	± 0.160 % to 0.006 % ± 0.006 % to 0.007 % ± 0.028 % to 0.030 % ± 0.030 % to 0.010 % ± 0.010 % to 0.008 %
6.	AC High Voltage (50 Hz)	WI-ET-02	1 kV to 28 kV	± 1.6 %
	DC High Voltage (Measure)		1 kV to 40 kV	± 1.6 %
7.	Resistance (Measure)	WI-ET-04	1 Ω to 100 Ω 100 Ω to 1 kΩ 1 kΩ to 10 kΩ 10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ	± 0.0120 % to 0.0012 % ± 0.0014 % to 0.0012 % ± 0.0012 % to 0.0013 % ± 0.0012 % to 0.0014 % ± 0.0015 % to 0.0030 % ± 0.0030 % to 0.0050 % ± 0.0050 % to 0.0130 %



Quality Manager

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: Peerless Calibration Services Ltd.
 2F, 8C & 9C, Zoo Road, Section-1, Mirpur, Dhaka-1216, Bangladesh.

Accreditation Standard: ISO/IEC 17025:2017 **Accreditation Date:** 29 October 2020
Certificate Number: 02.011.20 **Issued on:** 23 May 2023
Last Amended on: NA **Valid until:** 28 October 2026
Amendment no: NA

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
	Resistance (Source)	WI-ET-04	0.01 Ω to 1 kΩ 0.01 Ω to 1 MΩ 1 MΩ to 100 MΩ 0.1 GΩ to 1000 GΩ	± 0.0485 % to 0.0013 % ± 0.0450 % to 0.0013 % ± 0.0060 % to 0.0023 % ± 0.0060 % to 0.0023 %
8.	Frequency (Measure) Frequency (Source)	WI-ET-05 WI-ET-05	3 Hz to 500 kHz 45 Hz to 1 kHz	± 0.060 % to 0.006 % ± 0.12 % to 0.06 %
9.	Capacitance (Measure)	WI-ET-05	4 nF to 40 nF 40 nF to 400 nF 400 nF to 4 μF 4 μF to 40 μF 40 μF to 400 μF 100 μF to 1000 μF	± 0.12 %
	Capacitance (1 kHz) (Source)	WI-ET-05	1 nF to 100 μF	± 0.12 %
10.	Inductance (1 kHz) (Source)	WI-ET-05	100 μH to 1 mH 1 mH to 100 mH	± 0.31 % ± 0.12 %
11.	Watt (Measure)	WI-ET-05	Voltage 0 to 600 V Current 0 to 1000 Amp	± 0.31 %
Field: Thermal				
12.	Thermocouple / Temp. Controller (R-Type/ S-Type/ K-Type / E-Type/ J-Type / T-Type/ N-Type / B-Type/ L-Type / U-Type)	WI-TP-02	-200 °C to 100 °C 100 °C to 300 °C 300 °C to 800 °C 800 °C to 1200 °C 1200 °C to 1750 °C	± 0.08 to 0.1 °C ± 0.08 to 0.1 °C ± 0.08 to 0.1 °C ± 0.08 to 0.4 °C ± 0.08 to 0.8 °C
	RTD Sensor / Temperature Controller (PT-100/ PT-500 / PT-1000)	WI-TP-02	-200 °C to 800 °C	± 0.08 to 0.3 %


 Quality Manager

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: Peerless Calibration Services Ltd.
 2F, 8C & 9C, Zoo Road, Section-1, Mirpur, Dhaka-1216, Bangladesh.

Accreditation Standard: ISO/IEC 17025:2017 **Accreditation Date:** 29 October 2020
Certificate Number: 02.011.20 **Issued on:** 23 May 2023
Last Amended on: NA **Valid until:** 28 October 2026
Amendment no: NA

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
13.	Thermocouple / Temperature Controller (R-Type / S-Type/ K-Type / E-Type/ J-Type / T-Type/ N-Type / B-Type/ L-Type / U-Type	WI-TP-02	-200 °C to 100 °C 100 °C to 300 °C 300 °C to 800 °C 800 °C to 1200 °C 1200 °C to 1750 °C	± 0.08 to 0.1 °C ± 0.08 to 0.1 °C ± 0.08 to 0.1 °C ± 0.08 to 0.4 °C ± 0.08 to 0.8 °C
	RTD Sensor / Temperature Controller (PT- 100/ PT-500 / PT-1000)	WI-TP-02	-200 °C to 800 °C	± 0.08 to 0.3 %
14.	Glass Type Thermometer/ Digital Thermometer / Black Body Source / Hot Surface /Hot Plat/ Temperature Chamber/Incubator / Oven / Temperature Controller / Auto Clave / Refrigerator / Temperature Bath / Whirling Hygrometer	WI-TP-03	-40 °C to 0 °C 0 °C to 50 °C 50 °C to 100 °C 100 °C to 200 °C 200 °C to 400 °C 400 °C to 650 °C	± 0.1 °C ± 0.1 °C ± 0.1 °C ± 0.2 °C ± 0.3 °C ± 0.4 °C
	Muffle Furnace or Dry Oil Blocks	WI-TP-09	200 °C to 400 °C 400 °C to 1200 °C	± 0.8 °C ± 1.2 °C
	Infrared Thermometer / Non-Contact Pyrometers / Thermal Imager		0 °C to 300 °C 0 °C to 500 °C	± 1.2 °C ± 1.5 °C
15.	Humidity Chamber / Thermo-Hygrometer /	WI-TP-09	10 %RH to 95 %RH 10 to 80 °C	± 1.2 % ± 0.2 °C




 Quality Manager

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address:	Peerless Calibration Services Ltd. 2F, 8C & 9C, Zoo Road, Section-1, Mirpur, Dhaka-1216, Bangladesh.		
Accreditation Standard:	ISO/IEC 17025:2017	Accreditation Date:	29 October 2020
Certificate Number:	02.011.20	Issued on:	23 May 2023
Last Amended on:	NA	Valid until:	28 October 2026
Amendment no:	NA		

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
	Thermo-Hygrograph / Dehumidifier			
Field : Density				
16.	Density Hydrometer/ Viscosity Meter	WI-DS-01	0.6 g/ml to 2.0 g/ml	± 0.0006 to 0.0065 g/ml
Field: Force				
17.	Uniaxial Testing Machine (Universal, Tensile, Compression testing machine) & Load Cell with Indicator	WI-FC-01	1 cN to 1 kN 1 N to 1 kN 1 N to 5 kN 1 N to 10 kN 1 N to 20 kN 1 kN to 50 kN 1 kN to 100 kN	± 0.2 % ± 0.2 % ± 0.34 % ± 0.34 % ± 0.40 % ± 0.60 % ± 0.60 %
18.	Push Pull Tester / Force Gauge / Tension Meter	WI-FC-02	0 to 30 kgf 0 to 50 kgf Up to 500 gf	± 0.1 to 0.5 gf ± 0.1 to 0.8 gf ± 0.05 to 0.1 gf
Field: Pressure				
19.	Pressure Gauge (Pneumatic / Hydraulic), Pressure Indicating Device & Vacuum Gauge	WI-PS-01 WI-PS-02 WI-PS-03	-1000 to 2000 mbar 0 to 2 bar 0 bar to 30 bar 0 bar to 60 bar 0 bar 100 bar 0 bar to 1000 bar	± 0.012 to 0.018 mbar ± 0.0032 bar ± 0.0074 bar ± 0.034 bar ± 0.046 bar ± 0.144 bar
	Analog / Digital Barometer		300 hPa to 1100 hPa	± 1.9 hPa
Field: Hardness				
20.	Rubber Hardness Tester- (Type A, B, C, D, M, AO, DO, AM)	WI-HD-01	0 to 2.5 mm 0 to 100 Shore 0 to 50 N	± 0.005 mm ± 0.5 % ± 0.2 N
21.	Torque Meter, Torque Analyzer, Torque Driver, Torque Wrench	WI-HD-03	1.5 – 100 kgf.cm 0.15 - 10 N.m 10 to 1000 N.m	± 0.35 % ± 0.35 % ± 0.22 %
Field: Lux				



Quality Manager

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: Peerless Calibration Services Ltd.
2F, 8C & 9C, Zoo Road, Section-1, Mirpur, Dhaka-1216, Bangladesh.

Accreditation Standard: ISO/IEC 17025:2017 **Accreditation Date:** 29 October 2020
Certificate Number: 02.011.20 **Issued on:** 23 May 2023
Last Amended on: NA **Valid until:** 28 October 2026
Amendment no: NA

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
22.	Lux Meter / Light Box / Light Source / Pilling Viewer / Pili scope	WI-LX-01	10 Lux to 20000 Lux	± 2.5 % of rdg
Field: Speed				
23.	Non-Contact / Contact Tachometer / Travel Speed / Surface Speed / Amplitude / Elongation	WI-SP-01	5 RPM to 60000 RPM 0 mm to 2000 mm/min Up to 50 m/min	± 0.6 to 3.6 rpm ± 0.2 to 0.5 m/min
Field: Time				
24.	Digital/Analog Timer/ Stop Watches/Time Interval Meter / On-off Timer / Breaker Timer / Relay Timer	WI-ET-03	200 ms to 1 s 1 s to 60 s 1 m to 1 hr 1 hr to 4 hr 4 hr to 8 hr 8 hr to 16 hr 16 hr to 24 hr 24 hr to 48 hr 48 hr to 72 hr	± 0.053 sec ± 0.6 sec ± 0.0017 hr ± 0.0017 hr ± 0.0059 hr ± 0.0059 hr ± 0.0059 hr ± 0.0059 hr ± 0.0059 hr
Field: Sound				
25.	Acoustics Sound Level Meter / Sound Level Calibrator(1 kHz)	WI-SD-01	94 dB & 114 dB	± 0.6 dB
Field: Volume				
26.	Micro Pipette / Multichannel Micropipette / Auto Pipette / Micro Syringe / Auto Dispenser / Glass Pipette / Burette / Volumetric Flask / Measuring Cylinder / Beaker / Pycnometer	WI-VM-01 WI-VM-02 WI-VM-03	1 µl to 100 µl 100 µl to 500 µl 500 µl to 1000 µl 1 ml to 5 ml 5 ml to 10 ml 10 ml to 50 ml 50 ml to 100 ml 100 ml to 500 ml 500 ml to 1000 ml 1000 ml to 5000 ml 5000 ml to 20000 ml	± 0.1 µl ± 0.27 µl ± 0.41 µl ± 3 µl ± 5 µl ± 7 µl ± 8 µl ± 0.3 ml ± 0.6 ml ± 0.8 ml ± 2.5 ml



Quality Manager


SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: Peerless Calibration Services Ltd.
 2F, 8C & 9C, Zoo Road, Section-1, Mirpur, Dhaka-1216, Bangladesh.

Accreditation Standard: ISO/IEC 17025:2017 **Accreditation Date:** 29 October 2020
Certificate Number: 02.011.20 **Issued on:** 23 May 2023
Last Amended on: NA **Valid until:** 28 October 2026
Amendment no: NA

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
------	---------------------------------------	---------------------	-------------------------	--

Field:	Dimension			
27.	Dial / Digimatic / Vernier Caliper / Inside Dial Caliper	WI-DM-01	0 mm to 150 mm 0 mm to 300 mm 0 mm to 1000 mm	For L.C 0.01 mm ±10µm For L.C 0.05 mm ±11µm For L.C 0.1 mm ±15µm
	External Micrometer / Inside Micrometer / Outside Micrometer / Micrometer Head / Depth Micrometer	WI-DM-03	0 mm to 25 mm 25 mm to 100 mm 100 mm to 500 mm 500 mm to 1000 mm	For L.C 0.001 mm ±1µm For L.C 0.01 mm ±7µm For L.C 0.05 mm ±10µm For L.C 0.1 mm ±10µm
	Dial / Digimatic / Vernier Height Gauge	WI-DM-08	0 mm to 500 mm	For L.C 0.001 mm ± 10µm For L.C 0.01 mm ± 15µm
	Dial / Digital Thickness Gauge / Ultrasonic Thickness Gauge	WI-DM-05	Up to 100 mm	For L.C 0.001 mm ± 3µm For L.C 0.01 mm ± 5µm For L.C 0.01 mm ± 8µm
	Feeler Gauge / Foils/ Coating Thickness Gauge / Dial Indicator/ Lever Dial Gauge /Bore Gauge	WI-DM-04 WI-DM-15 WI-DM-16 WI-DM-18 WI-DM-12	0.005 mm to 2 mm 0.005 mm to 2 mm Up to 25 mm Up to 2 mm	± 2.6 µm ± 1.5 µm ± 7 µm ± 1 µm
	Master Rod / Length Bar / Extension Rod	WI-DM-09	Up to 500 mm	± 10 µm
	Scale	WI-DM-13	Up to 3000 mm	± 500√(L/1000) µm (L is in mm)



Quality Manager

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: Peerless Calibration Services Ltd.
 2F, 8C & 9C, Zoo Road, Section-1, Mirpur, Dhaka-1216, Bangladesh.

Accreditation Standard: ISO/IEC 17025:2017 **Accreditation Date:** 29 October 2020
Certificate Number: 02.011.20 **Issued on:** 23 May 2023
Last Amended on: NA **Valid until:** 28 October 2026
Amendment no: NA

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
	Tape / Pie Tape (Metallic, Woven, Fiber) / Circometer	WI-DM-14	Up to 50 m	$\pm 500\sqrt{L/1000} \mu\text{m}$ (L is in mm)
	Radius Gauge / Radius Template / Pitch Gauge / Ring Gauge / Plug Gauge / Pin Gauge / Snap Gauge / Go & No Go Gauge / Glass Scale / Slip Gauge Set / Test Sieves	WI-DM-22 WI-DM-23 WI-DM-18	10 μm to 40 mm	$\pm 1 \mu\text{m}$ to 15 μm
	Angle Gauge Block / Set Square / Triangle Set / Circular Protractor / Bevel Protractor / Angle Protractor	WI-DM-11	Up to 360 °	$\pm 5 \text{ min}$
	Microscope (Magnification)	WI-DM-19	Up to 200X	$\pm 0.30 \%$
	Acceleration / Velocity / Displacement	WI-ENV-03	0.1 to 199.99 m/s ² 0.1 to 199.99 mm/s 0.001 to 1.999 mm	± 0.20 to 4 %
Field:	Mass			
28.	Weights (Conventional Mass)	WI-MS-02	1- 5 mg 10 mg 20 mg 50 mg 100 mg	$\pm 0.006 \text{ mg}$ $\pm 0.008 \text{ mg}$ $\pm 0.010 \text{ mg}$ $\pm 0.012 \text{ mg}$ $\pm 0.016 \text{ mg}$




 Quality Manager

SCOPE OF ACCREDITATION (For Calibration Laboratory)

CAB Name & Address: Peerless Calibration Services Ltd.
 2F, 8C & 9C, Zoo Road, Section-1, Mirpur, Dhaka-1216, Bangladesh.

Accreditation Standard: ISO/IEC 17025:2017 **Accreditation Date:** 29 October 2020
Certificate Number: 02.011.20 **Issued on:** 23 May 2023
Last Amended on: NA **Valid until:** 28 October 2026
Amendment no: NA

S.N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2)
			200 mg	± 0.020 mg
			500 mg	± 0.025 mg
			1 g	± 0.030 mg
			2 g	± 0.040 mg
			5 g	± 0.050 mg
			10 g	± 0.060 mg
			20 g	± 0.080 mg
			50 g	± 0.10 mg
			100 g	± 0.16 mg
			200 g	± 0.30 mg
			500 g	± 0.80 mg
			1 kg	± 1.6 mg
			2 kg	± 3.0 mg
			5 kg	± 8.0 mg
			10 kg	± 50 mg
			20 kg	± 100 mg
			50 kg	± 250 mg
29.	Electronic Weighing Balance	WI-MS-01	1 mg to 100 g	For L.C 0.01mg ± 0.02mg
			100 g to 200 g	For L.C 0.1mg ± 0.2 mg
			200 g to 1000 g	For L.C 1mg ± 8 mg
			Up to 5 kg	For L.C 10 mg ± 0.5 g
			Up to 10 kg	For L.C 0.1 g ± 0.5 g
			Up to 30 kg	For L.C 1 g ± 1.2 g
			Up to 40 kg	For L.C 2 g ± 2.5 g
			Up to 500 kg	For L.C 5 g ± 5.5 g
			Up to 2000 kg	For L.C 5 g ± 18 g
Field: Flow				
30.	Water / Oil / Steam Flow	WI-FL-01	0 m ³ /hr to 12000 m ³ /hr	± 0.35 %
31.	Air Flow	WI-FL-02	0.1 m/s to 25 m/s	± 0.100 to 0.70 m/s
Field: Meter				
32.	PH Meter	WI-CH-01	PH -200 to 200mV & 2 to 16 pH	± 0.3 %
	Conductivity Meter	WI-CH-04	0 to 4000 µS/cm	± 0.6 %
33.	Moisture Meter	WI-ENV-02	0 to 90 %	± 0.5 %



Quality Manager



BANGLADESH ACCREDITATION BOARD (BAB)

91, Motijheel C/A, Dhaka-1000
Tel: +880-2-9513221
Fax: +880-2-9513222
Email: info@bab.gov.bd
Web: www.bab.gov.bd

Quality Manager