



# ACCREDITATION CERTIFICATE

Issued under the authority of Bangladesh Accreditation Act, 2006  
by Bangladesh Accreditation Board (BAB), Ministry of Industries to

**Cintex Co. UK Ltd.**

**House # 361 (2nd and 3rd Floor), Road # 5**

**Baitul Aman Housing Society, Adabor, Dhaka-1207, Bangladesh.**

This is to certify that this

## **Calibration Laboratory**

is accredited in accordance with the international standard  
**ISO/IEC 17025:2017**

in respect of the associated scope, subject to the terms and  
conditions governing the relevant conformity assessment  
body (CAB) accreditation.

Certificate Number : **02.021.23**  
Accreditation Date : **02 December 2023**  
Date of Issuance : **02 December 2023**  
Date of Expiration : **01 December 2026**



*Amam*  
**Sheikh Faezul Amin**  
Director General

This certificate must be returned on request; reproduction must follow BAB guidelines. For the specific scopes to which this accreditation applies, please refer to the Directory of CABs at BAB website.

**SCOPE OF ACCREDITATION**  
**(For Calibration Laboratory)**

**CAB Name & Address:** CINTEX CO. UK LTD., Calibration Laboratory

House # 361 (2<sup>nd</sup> and 3<sup>rd</sup> Floor), Road # 05, Baitul Aman Housing Society, Adabor, Dhaka-1207, Bangladesh

**Accreditation Standard:** ISO/IEC 17025:2017

**Accreditation Date:** 02 December 2023

**Certificate Number:** 02.021.23


**Issued on:** 02 December 2023

**Last Amended on:** NA

**Valid until:** 01 December 2026

**Amendment no:** NA

S. N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2) (±)
<b>Field: Mechanical (Balance)</b>				
01.	Mass Balance / Scale	CINTEX-SOP-02	0.01mg up to 220g	± 0.02 g
	Semi-Micro Balances		0.01mg up to 620g	± 0.02 g
	Analytical Balances		0.10 g up to 6 kg	± 0.26 g
	Top Loading Balances (In Lab & On Site)		5.00 g up to 30 kg	± 0.02 g
<b>Field: Mechanical (Weight)</b>				
02.	Mass (Weight) (In Lab & On Site)	CINTEX-SOP-01	1 mg	± 0.17 mg
			2 mg	± 0.17 mg
			5 mg	± 0.17 mg
			10 mg	± 0.17 mg
			20 mg	± 0.17 mg
			50 mg	± 0.17 mg
			100 mg	± 0.36 mg
			200 mg	± 0.22 mg
			500 mg	± 0.23 mg
			1 g	± 0.17 mg
			2 g	± 0.24 mg
			5 g	± 0.42 mg
			10 g	± 0.32 mg
			20 g	± 0.27 mg
			50 g	± 0.32 mg
			100 g	± 0.82 mg

  
 Quality Manager

## SCOPE OF ACCREDITATION (For Calibration Laboratory)

**CAB Name & Address:** CINTEX CO. UK LTD., Calibration Laboratory

House # 361 (2<sup>nd</sup> and 3<sup>rd</sup> Floor), Road # 05, Baitul Aman Housing Society, Adabor,  
Dhaka-1207, Bangladesh

**Accreditation Standard:** ISO/IEC 17025:2017

**Accreditation Date:** 02 December 2023

**Certificate Number:** 02.021.23

**Issued on:** 02 December 2023


**Last Amended on:** NA

**Valid until:** 01 December 2026

**Amendment no:** NA

S. N.	Measured quantity Instrument/Gauge	Reference to Method	Measurement range/value	Measurement Capabilities expressed as expanded uncertainty U (k=2) (±)
			200 g	± 0.51 mg
			500 g	± 4.03 mg
			1 kg	± 0.16 g
			2 kg	± 0.16 g
			5 kg	± 0.24 g
			10 kg	± 2.16 g
			20 kg	± 1.34 g
<b>Field: Mechanical (Length)</b>				
03.	Length (Steel Scale) (In Lab & On Site)	CINTEX-SOP-03	10 mm	±0.03 mm
			50 mm	±0.01 mm
			100 mm	±0.01 mm
			500 mm	±0.03 mm
			1000 mm	±0.05 mm

-----End-----

  
 \_\_\_\_\_  
 Quality Manager